



**Creatinine** measurements are used as an aid to monitor and diagnose **renal disease**.

METHOD: MODIFIED JAFFÉ; KINETIC

**FEATURES:**

FOR THE QUANTITATIVE MEASUREMENT OF CREATININE  
**Genzyme Creatinine-S Reagent**

Genzyme's Creatinine-S method is a modified kinetic, Jaffé procedure, capable of further reducing potential interferences when compared to conventional kinetic creatinine methods. It is intended for the measurement of creatinine concentration in serum and urine.

- ✓ Two part stable liquid, ready to use reagent
- ✓ No significant lipemic interference
- ✓ Applicable to multiple chemistry platforms
- ✓ Serum and urine samples can be used

**BENEFITS:**

- ✓ Easy to use, no additional reagent preparation required
- ✓ Minimizes the need for sample dilutions
- ✓ High reliability of testing
- ✓ Laboratory flexibility

ORDERING INFORMATION	CONFIGURATION	CATALOG NUMBER
<input type="radio"/> GENZYME CREATININE-S	R1 2 X 250 mL R2 1 X 125 mL 1 X 15 mL CALIBRATOR*	221-30
<input type="radio"/> GENZYME CREATININE-S	R1 1 X 1000 mL R2 1 X 250 mL	221-50
<input type="radio"/> GENZYME DC-CAL CALIBRATOR	5 X 3 mL	SE-035
<input type="radio"/> GENZYME DC-TROL LEVEL 1	10 X 5 mL	SM-052
<input type="radio"/> GENZYME DC-TROL LEVEL 2	10 X 5 mL	SM-056

\*Calibrator for Manual Assays Only



**THE AMERICAS**  
 31 New York Avenue  
 Framingham, MA 01701-9322  
 Phone: 800 332 1042  
 Fax: 800 762 6311  
 Email: usdiagcustomerservice@genzyme.com

**INTERNATIONAL**  
 50 Gibson Drive  
 Kings Hill, West Malling  
 Kent, ME19 4AF, UK  
 Phone: +44 (0) 1732 220022  
 Fax: +44 (0) 1732 220024/5  
 Email: ukdiagcustomerservice@genzyme.com

