



Triglyceride measurements are used as an aid to monitor and diagnose **heart disease**.

METHOD: ENZYMATIC (GLYCEROL PHOSPHATE OXIDASE); ENDPOINT

FEATURES:

FOR THE QUANTITATIVE MEASUREMENT OF TRIGLYCERIDE

Genzyme Triglyceride-SL Reagent

Genzyme's Triglyceride-SL Reagent uses a fully enzymatic single reagent procedure intended for the measurement of triglyceride concentration in serum.

- ✓ One part stable liquid, ready to use reagent
- ✓ Wide Linear Range
- ✓ Applicable to multiple instrument platforms

BENEFITS:

- ✓ Saves time and labor while reducing errors and reagent waste
- ✓ Reduces repeat testing
- ✓ Laboratory flexibility

ORDERING INFORMATION	CONFIGURATION	CATALOG NUMBER
<input type="radio"/> GENZYME TRIGLYCERIDE-SL	R1 5 X 100 mL	236-17 <small>(ROCHE/HITACHI® COMPATIBLE WEDGE)</small>
<input type="radio"/> GENZYME TRIGLYCERIDE-SL	R1 2 X 100 mL	236-60
<input type="radio"/> GENZYME TRIGLYCERIDE-SL	R1 1 X 1000 mL	236-99
<input type="radio"/> GENZYME DC-CAL	5 X 3 mL	SE-035
<input type="radio"/> GENZYME DC-TROL LEVEL 1	10 X 5 mL	SM-052
<input type="radio"/> GENZYME DC-TROL LEVEL 2	10 X 5 mL	SM-056



THE AMERICAS
 31 New York Avenue
 Framingham, MA 01701-9322
 Phone: 800 332 1042
 Fax: 800 762 6311
 Email: usdiagcustomerservice@genzyme.com

INTERNATIONAL
 50 Gibson Drive
 Kings Hill, West Malling
 Kent, ME19 4AF, UK
 Phone: +44 (0) 1732 220022
 Fax: +44 (0) 1732 220024/5
 Email: ukdiagcustomerservice@genzyme.com

