CLINICAL CHEMISTRY REAGENTS

Acetaminophen L3K®

FOR THE QUANTITATIVE MEASUREMENT OF ACETAMINOPHEN

METHOD: ENZYMATIC (ACYL AMIDOHYDROLASE); ENDPOINT

Acetaminophen (paracetamol) measurements are used in the diagnosis and treatment of acetaminophen overdose toxicity.

Sekisui's Acetaminophen L3K® assay uses a spectrophotometric method that is rapid, reliable, convenient and specific for acetaminophen. It is intended for use with serum and plasma.

Features: • Two part stable liquid Reduced NAC interference • A single level liquid acetaminophen Suitable for use with serum and calibrator is included in the kit lithium heparin plasma **Benefits:** Easy to use, no additional reagent Confidence in results preparation required Flexible sample types to meet Convenient and cost effective different laboratory needs Performance Characteristics **Precision No Significant Interferences**

- Within-Run: ≤1.5%
- Total Precision: ≤2.9%

Accuracy

SERUM^(a)

- Slope: 1.064
- Intercept: 1.1 μ g/mL (7.0 μ mol/L)
- Correlation Coefficient: 0.9998

PLASMA^(b)

- Slope: 0.999
- Intercept: -0.3 μ g/mL (-2.2 μ mol/L)
- Correlation Coefficient: 0.9999

Linearity

- 0.6-377.5 μg/mL (4-2500 μmol/L)
- (a) The performance of this method (y) was compared with the performance with a similar acetaminophen method (x) on a Roche/Hitachi® 717 analyzer.
- (b) The performance of this method with plasma (y) was compared with the performance of this method with serum (x) on a Roche/Hitachi[®] 717 analyzer.

Up to Levels Indicated

- Hemoglobin: 200 mg/dL (31 µmol/L)
- Conjugated Bilirubin: 2 mg/dL (23.7 µmol/L)
- Unconjugated Bilirubin: 2 mg/dL (34.2 µmol/L)
- Ascorbic Acid: 3000 µg/dL (170 µmol/L)
- N-Acetylcysteine: 1500 mg/L (9.2 mmol/L)
- Intralipid: 200 mg/dL (600 mg/dL) (6.8 mmol/L) Simulated Triglycerides)

Reference Range⁽¹⁾

- Therapeutic Concentration: 10-30 µg/mL (66-199 µmol/L)
- Toxic Concentration: >200 µg/mL (1324 µmol/L)

CLINICAL CHEMISTRY REAGENTS

Acetaminophen L3K®

FOR THE QUANTITATIVE MEASUREMENT OF ACETAMINOPHEN

METHOD: ENZYMATIC (ACYL AMIDOHYDROLASE); ENDPOINT

Acetaminophen (paracetamol) measurements are used in the diagnosis and treatment of acetaminophen overdose toxicity.

Sekisui's Acetaminophen L3K[®] assay uses a spectrophotometric method that is rapid, reliable, convenient and specific for acetaminophen. It is intended for use with serum and plasma.

Ordering information		
	Configuration	Catalog Number
Acetaminophen L3K®	R1 1 x 10 mL R2 2 x 10 mL 1 x 5 mL (Calibrator)	506-10
Acetaminophen L3K®	R1 3 x 10 mL R2 6 x 10 mL 1 x 5 mL (Calibrator)	506-30

(1) Burtis, C.A. and Ashwood, E.R., Eds, Tietz Tetbook of Clinical Chemistry, Second Edition, pp 1168, 2212, W.B. Saunders Company, Philadelphia (1994).



Experience + Technology + Portfolio + Support = CHEMABILITY

THE AMERICAS

Sekisui Diagnostics, LLC One Wall Street Burlington, MA 01803 Phone: 800 332 1042 Fax: 800 762 6311 Email: questions@sekisui-dx.com

INTERNATIONAL

Sekisui Diagnostics (UK) Limited Liphook Way, Allington Maidstone, Kent, ME16 0LQ, UK Phone: +44 1622 607800 Fax: +44 1622 607801 Email: info@sekisui-dx.com



Because every result matters™

www.sekisuidiagnostics.com

© 2020 Sekisui Diagnostics, LLC. All rights reserved. SEKURE® and Chemistry Care® are registered trademarks of Sekisui Diagnostics, LLC. SK[™] logo, ChemAbility™, and *Because every result matters*™ are trademarks of Sekisui Diagnostics, LLC.